

# i-Tracker<sup>®</sup>

## anti-Adalimumab


English

REF

CTA 003-100



### INTENDED USE

**i-Tracker anti-Adalimumab** () is an automated assay intended for the quantitative measurement of Anti-Adalimumab antibodies (anti-drug antibodies: ADA<sub>b</sub>) in human serum or plasma samples.


### DIAGNOSTIC VALUE

Anti-TNF $\alpha$  are therapeutic agents widely used to treat patients with various inflammatory diseases. Adalimumab is one of the anti-TNF $\alpha$  recommended for the treatment of the rheumatoid arthritis, Crohn's disease, ankylosing spondylitis, etc. This human monoclonal antibody is able to bind TNF $\alpha$ . It blocks the action of TNF $\alpha$  responsible for the inflammatory state.

However, during the treatment, some patients can develop antibodies against Adalimumab. Consequently, the plasmatic level of anti-TNF $\alpha$  decreases and simultaneously the disease symptoms reappear or increase.

Studies have shown that the trough level of a TNF $\alpha$  therapy (i.e. the circulating drug level just before the next injection) usually correlates clinical efficacy. This trough level is influenced by several factors, among them dosage and frequency of the injections, disease phenotype and activity, pharmacogenetic factors, co-medication and the formation of anti-drug antibodies.

Measurement of anti-TNF $\alpha$  drug level in combination with anti-drug antibodies quantification provides to the treating physician help for therapeutic guidance, hence maximizing treatment efficacy while minimizing to major cost savings.

**i-Tracker anti-Adalimumab** () assay is validated to monitor anti-drug antibodies induced by any biological drug which contains the active substance Adalimumab, that is the original drug Humira<sup>®</sup>, and any biosimilar drug like ABP501 (Amjevita<sup>®</sup>) and SB5 (Imraldi<sup>®</sup>).

### SPECIMEN COLLECTION

The specimen types appropriate to **i-Tracker anti-Adalimumab** are human sera and plasma.

Samples which are cloudy should be clarified by low-speed centrifugation.

To prevent erroneous results due to the presence of fibrin, ensure that complete clot formation has taken place prior to centrifugation of samples. Some samples, particularly those from patients receiving anticoagulant therapy, may require increased clotting time.

Freshly collected specimens could be tested after storage of at most 8 days kept in refrigerator (+2°C / +8°C) or 3 days at room temperature (+18°C / +25°C) and until 10 hours for onboard specimens.

5 freeze-thaw cycles for specimens do not affect the testing results.

### METHOD PRINCIPLE

**i-Tracker anti-Adalimumab** assay is a two-step immunoassay using microparticles, acridinium-ester labeled chemiluminescent technology with the i-Track<sup>10</sup>.

- In the first step, the Adalimumab coupled magnetic microparticles, and human serum/plasma sample are mixed in an assay cuvette, which allows Anti-Adalimumab antibodies to bind to the microparticles surface.
- After incubation, unbound reagent and sample matrix are removed by washing, and the microparticles-Adalimumab-Anti-Adalimumab antibodies immunocomplex are kept with the help of a magnetic separator.
- Secondly, Adalimumab conjugated to acridinium ester is added.
- After incubation, excess acridinium-ester conjugate is removed by washing and finally the light induced by acridinium-ester is detected by addition of triggers.
- The relative light unit (RLU) intensity is proportional to the amount of Anti-Adalimumab antibodies.
- According to a certain specific Anti-Adalimumab antibodies RLU-concentration standard curve, the RLU obtained can be interpreted to Anti-Adalimumab antibodies concentration in the sample expressed as ng/mL.

For quantitation of Anti-Adalimumab antibodies, the **i-Tracker Anti-Adalimumab** assay uses a predefined lot specific Master Curve that is uploaded into the instrument through the reagent cartridge 2D barcode. The Master Curve is created during manufacturing by using in-house standards. Based on the Master Curve, and results obtained by running two Calibrators, an instrument specific Working Curve is created, which is used to calculate a concentration (ng/mL) from the RLU obtained for each sample.

## KIT CONTENTS

Components are a matched set. **Barcode on the inside box are needed for the assay.**

Description	Volume
<b>Cartridge reagents</b>	
<p><b>Microparticles</b> : 1 vial of Adalimumab coupled microparticles in PBS with stabilizers. <span style="border: 1px solid black; padding: 2px;">MP</span> <span style="border: 1px solid black; padding: 2px;">ADAA</span></p> <p>Preservatives: Sodium azide &lt; 0.1% and Proclin 300 &lt; 0.06%</p> <p><b>Ready to use</b></p>	3 mL*
<p><b>anti-Adalimumab Buffer</b> : 1 vial of PBS-Tween with stabilizer. <span style="border: 1px solid black; padding: 2px;">BUF</span> <span style="border: 1px solid black; padding: 2px;">ADAA</span></p> <p>Preservative: Proclin 300 &lt;0.06%</p> <p><b>Ready to use</b></p>	23 mL*
<p><b>Immunoconjugate</b>: 1 vial of acridinium-ester labeled Adalimumab, containing PBS with stabilizers. <span style="border: 1px solid black; padding: 2px;">IC</span> <span style="border: 1px solid black; padding: 2px;">ADAA</span></p> <p>Preservatives: Sodium azide &lt; 0.1% and Proclin 300 &lt; 0.06%</p> <p><b>Ready to use</b></p>	23 mL*
<b>Control and Calibrators</b>	
<p><b>anti-Adalimumab Calibrator A</b>: a calibrator tube containing Anti-Adalimumab antibodies (low level) in stabilizers and preservatives. <span style="border: 1px solid black; padding: 2px;">CALA</span> <span style="border: 1px solid black; padding: 2px;">ADAA</span></p> <p><b>Target value for the calibrator is indicated on the 2D barcode localized in each kit.</b></p> <p>Preservatives: Sodium azide &lt; 0.1% and Proclin 300 &lt; 0.06%</p> <p><b>Ready to use</b></p>	2 mL*
<p><b>anti-Adalimumab Calibrator B</b> : a calibrator tube containing Anti-Adalimumab antibodies (high level) in stabilizers and preservatives. <span style="border: 1px solid black; padding: 2px;">CALB</span> <span style="border: 1px solid black; padding: 2px;">ADAA</span></p> <p><b>Target value for the calibrator is indicated on the 2D barcode localized in each kit.</b></p> <p>Preservatives: Sodium azide &lt; 0.1% and Proclin 300 &lt; 0.06%</p> <p><b>Ready to use</b></p>	2 mL*
<p><b>anti-Adalimumab Control</b>: a control tube containing Anti-Adalimumab antibodies in stabilizers and preservatives. <span style="border: 1px solid black; padding: 2px;">CTL1</span> <span style="border: 1px solid black; padding: 2px;">ADAA</span></p> <p><b>Acceptable range for the control is indicated on the 2D barcode localized in each kit.</b></p> <p>Preservatives: Sodium azide &lt; 0.1% and Proclin 300 &lt; 0.06%</p> <p><b>Ready to use</b></p>	2 x 2 mL*

\*When loading a reagent, the i-Track<sup>10</sup> system indicates the remaining usable volume (dead volume is not taken into account).

## MATERIALS REQUIRED BUT NOT PROVIDED

- i-Track<sup>10</sup> (Cat. No. TD 810400)
- Cartridge Checking System (CCS) (Cat. No. TD IS-6010)
- Cuvettes (Cat. No. TD IS-CC100)
- System Liquid (Syst.L) (Cat. No. TD IS-CS100)
- Trigger Set (Cat. No. TD IS-CT100)
- Wash Solution (Wash S) (Cat. No. TD CW100)
- Immunocleaner, (Cat. No. TD IS-IM100)
- D-Sorb Solution (Cat. No. TD IS-DS200)
- Disposable Waste Bags, (Cat No. TD IS-DW225)
- Barcode scanner
- XPrep


## STABILITY AND STORAGE CONDITIONS

- The kit is stable until the expiration date when stored and handled as directed.
- Store the kit in refrigerator (+2°C / +8°C). Open reagents can be kept on board for a maximum of 30 days, or stored at +2°C / +8°C for a maximum of 60 days. When one of the two conditions is reached the reagents are no longer usable.
- The i-Track<sup>10</sup> software monitors the onboard (in-use) expiration of the reagent cartridge. The system will not allow use of a reagent which has expired.

## PRECAUTIONS

1. The product is for in vitro diagnostic use only.
2. This assay is only for use on the i-Track<sup>10</sup>.
3. Do not use the reagents beyond their expiration dates. Do not mix reagents from different lots.
4. Instructions must be carefully followed for reagent use and storage. Any modification in procedure may interfere with the results. The control, calibrators and contaminated vials must strictly follow safety guidelines or rules of biological hazards to ensure the users and environmental safety.
5. The human derived material in this product was tested and found nonreactive for Hepatitis B Surface Antigen (HBsAg), Anti-HCV and HIV 1/2 antibodies. Handle as if potentially infectious. Avoid contact with skin and eyes. Do not empty into drains. Wear suitable protective clothing.
6. This product requires the handling of control, calibrators and human specimens which contain human sourced materials. It is recommended that all human sourced materials be considered potentially infectious and handled in accordance with the OSHA Standard on Bloodborne Pathogens. Biosafety Level 2 or other appropriate biosafety practices should be used for materials that contain or are suspected of containing infectious agents.
7. Reagents contain chemical and biological components. Avoid ingesting or splashing onto skin and mucous membrane. If direct contact with controls happens, rinse

immediately the contact area with plenty of water and see a doctor if necessary.

8. Liquid waste and solid waste are temporarily stored on the i-Track<sup>10</sup> in separate containers. Waste management should also be handled in accordance with standards mention in the chapter Warnings and Precautions point No. 6.
9. Spilled reagents should be cleaned up immediately. Observe all federal, state and local environmental regulations when disposing of wastes.
10. Once opened, the reagent cartridge must be stored in the instrument's refrigerated reagent racks. Care should be taken to avoid spilling the reagents when the reagent cartridge is reloaded into the instrument.
11. Chemical contamination of the reagents can result from improper cleaning or rinsing of the instrument. Residues from common laboratory chemicals such as formalin, bleach, ethanol, or detergent can cause interference in the assay. Be sure to follow the recommended cleaning procedure of the instrument as outlined in the i-Track<sup>10</sup> user manual.
12. Reagents can contain Sodium azide < 0.1% and/or Proclin 300 <0.06%. Do not eat and avoid contact with skin and eyes. Azide can form explosive mixtures in copper or lead piping. Rinse thoroughly after flushing.
13.  At this concentration, ProClin 300 is irritating to eyes and skin, and may be detrimental if enough quantity is ingested. It is a skin sensitizer; prolonged or repeated exposure may cause allergic reaction in certain sensitive individuals.
14. Avoid using reagents if signs of contamination or other visible changes occur.

## ASSAY PROCEDURE

Place the reagent cartridge in the reagent compartment making sure that the barcode is read.

The calibrators and the control must be placed in the sample compartment and make sure that the barcode is read.

Select the appropriate assay protocol on the i-Track<sup>10</sup>.

Note that for optimal performance, it is important to perform all routine maintenance procedures, such as routine cleaning, calibration and control procedures which are defined in the i-Track<sup>10</sup> User Manual.

See the i-Track<sup>10</sup> User Manual for preparation, setup, dilutions, adjustment, assay and quality control procedures.

Users should have the periodic calibration procedure for every 21 running days from last calibration, which will be reminded on the software interface. Besides, a calibration procedure should be carried out when a new batch of **i-Tracker anti-Adalimumab** kit is used.

The control procedure could be done before running the specimens every day. Users also can adjust the control procedure period according to their own lab frequency.

## ASSAY CHARACTERISTICS & PERFORMANCES

### Lower Limit of Quantification (LLOQ) determination

A population of serum samples from healthy donors or untreated patients were valued by **i-Tracker anti-Adalimumab** assay.

Lower Limit of Quantification of anti-Adalimumab
10 ng/mL > 95 <sup>th</sup> percentile

### Dynamic range & Test Result Interpretation

With the help of the predicated two-point calibration master curve and a working curve for the instrument specified, the i-Track<sup>10</sup> will automatically calculate the Anti-Adalimumab antibodies concentration of each specimen and interpret the results into ng/mL.


Measurement range anti-Adalimumab
10 ng/mL - 2000 ng/mL

Results below the lower limit (<10 ng/mL) will be reported as HDM-, while those above the upper limit (>2000 ng/mL) will be reported as HDM+.

Specimen with concentration < 10 ng/mL is unquantifiable.


Test results only reflect the sample collecting status and should be interpreted/analyzed for diagnosis in conjunction with other laboratory and clinical findings.

### Interfering Substances studies

**i-Tracker anti-Adalimumab** () assay was evaluated to assess the impact of potential interfering molecules (bilirubin (0.2 mg/mL), hemoglobin (2 mg/mL), triglycerides (10 mg/mL), rheumatoid factors (1000 IU/mL) and biotin (2000 ng/mL)).

⇒ No interference is detected.

### Cross Reactivity

**i-Tracker anti-Adalimumab** () assay was evaluated to assess the impact of potential cross reacting molecules (Infliximab, Certolizumab, Etanercept, Golimumab, Ustekinumab, Vedolizumab and anti-Infliximab antibodies).


⇒ No cross reactivity is detected.

### Precision

Parameter	Intra-run (10 tests in a same assay)		Inter-run (2 tests in 6 independent assay)	
	Mean (ng/mL)	CV (%)	Mean (ng/mL)	CV (%)
Anti-Adalimumab	31	3.1	36	1.7
	251	3.0	300	0.8
	1328	2.6	1406	3.6

### **Correlation :**

The quantification of Anti-Adalimumab antibodies performed with **i-Tracker Anti-Adalimumab** assay gave similar results than the quantification made with **LISA-TRACKER Anti-Adalimumab** assay (Cat.n: LTA 003/LTA 005). Evaluation performed with 46 samples, shown that both assays are equivalent (Spearman's coefficient = 0.97).



### **LIMITATIONS**

The effectiveness of this kit is only confirmed for human sera/plasma, the applicability of the other kinds of samples is not verified.

Reliable and reproducible results will be obtained when the assay procedure is carried out in accordance with the instructions and with adherence to good laboratory practice

Clinical diagnosis should not be made on the findings of a single test result but should integrate all clinical and laboratory findings.

### **QUALITY CONTROL**

It is recommended to use internally and externally sourced control material. **IMMUNO-TROL i-Tracker anti-Adalimumab** control (<sup>®</sup>Theradag), (Cat.n: CTA 003-PC) contains **anti-Adalimumab antibodies**. This material is to be assayed in the same manner as the unknown sample.

### **REFERENCES**

**Arends S & al.** The formation of autoantibodies and antibodies to TNF- $\alpha$  blocking agents in relation to clinical response in patients with ankylosing spondylitis. *Clin Exp Rheumatol.* 2010 Sep-Oct;28(5):661-8.

**Bartelds GM & al.** Development of antidrug antibodies against adalimumab and association with disease activity and treatment failure during long-term follow-up. *JAMA.* 2011 Apr 13;305(14):1460-8.

**Bendtzen K.** Is There a Need for Immunopharmacologic Guidance of Anti-Tumor Necrosis Factor Therapies. *ARTHRITIS & RHEUMATISM* Vol. 63, No. 4, April 2011, pp 867–870.

**Bressler B & al.** Clinical Practice Guidelines for the Medical Management of Nonhospitalized Ulcerative Colitis: The Toronto Consensus. *Gastroenterology* 2015;148:1035 – 1058.

**Choon Jin Ooi & al.** Best practices on immunomodulators and biologic agents for ulcerative colitis and Crohn's disease in Asia. *Intest Res,* May 31, 2019:1-26.

**Choy & al.** Efficacy of a novel PEGylated humanized anti-TNF (CDP870) in patients with rheumatoid arthritis : a phase II double-blinded, randomized, dose escalating trial. *Rheumatology* 2002;41:1133-1137.

**Clinical Guideline:** Management of Crohn's Disease in Adults. *Am J Gastroenterol* 2018; 113:481–517.

**Desroches M & al.** Treatment failure with antagonists of TNF- $\alpha$ : mechanisms and implications for the care of patients. *Eur. Cytokine Netw.,* Vol. 21 n° 4, December 2010, 226-31.

**Diagnostics Guidance [DG22] on Therapeutic monitoring of TNF-alpha Inhibitors in Crohn's Disease.**  
<https://www.nice.org.uk/guidance/dg22/chapter/1-Recommendations>

**ECCO-ESGAR Guideline for Diagnostic Assessment in IBD Part 1:** Initial diagnosis, monitoring of known IBD, detection of complications. *Journal of Crohn's and Colitis,* 2018, 1–32.

**Feuerstein J D. & al.** Therapeutic Drug Monitoring in Inflammatory Bowel Disease. *Gastroenterology.* 2017:1-8.

**Gomollón F & al.** 3rd European Evidence-based Consensus on the Diagnosis and Management of Crohn's Disease. *J Crohns Colitis.* 2016:1-23.

**Greuter T & al.** Therapeutic Drug Monitoring to Guide Clinical Decision Making in Inflammatory Bowel Disease Patients with Loss of Response to Anti-TNF: A Delphi Technique-Based Consensus. *Digestion.* 2019 Aug 28:1-9.

**Grossi V & al.** Anti-adalimumab and anti-certolizumab antibodies titers after discontinuation of adalimumab: two case reports. *Clin Chem Lab Med.* 2019 Oct 30.

**Jaminitski & al.** The presence or absence of antibodies to infliximab or adalimumab determines the outcome of switching to etanercept. *Rheum Dis.* 2011 Feb;70(2):284-8.

**Khan A & al.** New Zealand Society of Gastroenterology Guidelines on Therapeutic Drug Monitoring in Inflammatory Bowel Disease. *N Z Med J.* 2019 Mar 8;132(1491):46-62.

**Koren & al.** Recommendation on risk-based strategies for detection and characterization of antibodies against biotechnology products. *Journal of Immunological Methods,* 333 (2008) 1-9.

**Korswagen LA & al.** Venous and Arterial Thromboembolic Events in Adalimumab-Treated Patients With Anti-adalimumab Antibodies. *ARTHRITIS & RHEUMATISM* Vol. 63, No. 4, April 2011, pp 877–883.

**Labetoulle R & al.** Prolonged Persistence of Adalimumab Transferred From Mother to Infant During Pregnancy. *Ann Intern Med.* 2018 Mar 6.

**Lamb CA & al.** British Society of Gastroenterology consensus guidelines on the management of inflammatory bowel disease in adults. *Gut* 2019;0:1–106.

**Little RD & al.** Comparison of adalimumab serum drug levels when delivered by pen versus syringe in patients with inflammatory bowel disease. An international, multicentre cohort analysis. *J Crohns Colitis.* 2019 May 15.

**Mahil SK & al.** Predicting treatment response in psoriasis using serum levels of adalimumab and etanercept: a single-centre, cohort study. *Br J Dermatol.* 2013 Aug;169(2):306-13.

**Management of Paediatric Ulcerative Colitis, Part 1: Ambulatory Care—An Evidence-based Guideline.** *J Pediatr Gastroenterol Nutr.* 2018 May 30.

**Marcus Harbord & al.** Third European Evidence-based Consensus on Diagnosis and Management of Ulcerative Colitis. Part 2: Current Management. *Journal of Crohn's and Colitis,* 2017, 1–24.

**Martelli L & al.** Cost-effectiveness of drug monitoring of anti-TNF therapy in inflammatory bowel disease and rheumatoid arthritis: a systematic review. *J Gastroenterol.* 2017 Jan;52(1):19-25.

**Mire-Sluis & al.** Recommendation for the design and optimization of immunoassays used in the detection of host antibodies against biotechnology products. *Journal of Immunological Methods,* 289 (2004) 1-16.

**Mitrev N & al.** Consensus statements on therapeutic drug monitoring of anti-tumour necrosis factor therapy in inflammatory bowel diseases. *Aliment Pharmacol Ther.* 2017 Dec;46(11-12):1037-1053.

**Papamichael K & al.** Appropriate Therapeutic Drug Monitoring of Biologic Agents for Patients With Inflammatory Bowel Diseases. *Clin Gastroenterology and Hepatology,* 24 March 2019.

**Peyrin-Biroulet L & al.** French National Consensus Clinical Guidelines for the Management of Crohn's disease. *Dig Liver Dis.* 2016.

**Peyrin-Biroulet L & al.** French National Consensus Clinical Guidelines for the Management of Ulcerative Colitis. *Dig Liver Dis.* 2016; 48(7):726-33.

**Quistrebert J & al.** Incidence and risk factors for adalimumab and infliximab anti-drug antibodies in rheumatoid arthritis: A European retrospective multicohort analysis. *Semin Arthritis Rheum.* 2018 Oct 12. pii: S0049-0172(18)30176-8.

**Radstake TRDJ & al.** Formation of antibodies against infliximab and adalimumab strongly correlates with functional drug levels and clinical responses in rheumatoid arthritis. *Ann Rheum Dis* 2009;68:1739–1745.

**Ram Raj Singh, & al.** TNF $\alpha$  blockade in human diseases : mechanisms and future. *Clin.Immunol.,* 2008 February ; 126(2):121-136.

**Roblin X & al.** Association between pharmacokinetics of adalimumab and mucosal healing in patients with inflammatory bowel diseases. *Clin Gastroenterol Hepatol.* 2014 Jan;12(1):80-84.e2.

**Roblin X & al.** Development of an algorithm incorporating pharmacokinetics of adalimumab in inflammatory bowel diseases. *Am J Gastroenterol.* 2014 Aug;109(8):1250-6.

**Ruemmele F.M. & al.** Consensus guidelines of ECCO/ESPGHAN on the medical management of pediatric Crohn's disease. *ournal of Crohn's and Colitis* (2014).

**Sparado A & al.** Switching from infliximab or etanercept to adalimumab in resistant or intolerant patients with spondyloarthritis: a 4-year study. *Rheumatology (Oxford).* 2010 Jun;49(6):1107-11.

**Therapeutic Drug Monitoring in Inflammatory Bowel Disease.** *Gastroenterology.* 2017;153(3):835-857.

**Therapeutic Drug Monitoring in Inflammatory Bowel Disease.** *Gastroenterology.* 2017;153:858-859.

**3rd European Evidence-based Consensus on the Diagnosis and Management of Crohn's Disease.** *J Crohns Colitis.* 2017; 11(1):3-25.

**3rd European Evidence-based Consensus on the Diagnosis and Management of Ulcerative Colitis.** *J Crohns Colitis.* 2017; 11(6):649-670.

**Ungar B & al.** Addition of an immunomodulator can reverse antibody formation and loss of response in patients treated with adalimumab. *Aliment Pharmacol Ther.* 2017 Jan;45(2):276-282.

**Ward MG & al.** Infliximab and adalimumab drug levels in Crohn's disease: contrasting associations with disease activity and influencing factors. *Aliment Pharmacol Ther.* 2017 Jul;46(2):150-161.



## SYMBOLS USED



EC Declaration of conformity



*In Vitro* Diagnostic Medical Device



Catalogue number



Manufacturer



Lot Number



Consult Instruction For Use



Expiry Date



Temperature limitation



Number of tests



Biological hazard



Warning



14 rue Ambroise Croizat  
77183 CROISSY-BEAUBOURG  
France

Tel : +33 (0)1 64 62 10 12  
Fax : +33 (0)1 64 62 09 66

E-mail : [support@theradiag.com](mailto:support@theradiag.com)  
Internet : [www.theradiag.com](http://www.theradiag.com)